



广微测
Gmicro Testing



201819000883



中国认可
国际互认
检测
TESTING
CNAS L1747

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS

Report №.

[REDACTED]

Name of Sample

Sterilizer

Applicant

[REDACTED]

Test Type

Entrustment Test



Address: Building 66, No.100 Central Xian Lie Road, Guangzhou, China

Postcode: 510070

Tel: +86 20 87137666

Fax : +86 20 87137668

Website : www.gddcm.com



广微测
Gmicro Testing



201819000883



中国认可
国际互认
检测
TESTING
CNAS L1747

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS

Report No.:

Verification Code:

Name of Sample	Sterilizer	Test Type	Entrustment Test
Applicant		Address	
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	1PCS
Spec and Lot No of Sample	AST-WR125	State and Characteristic	Machine
Sample Received Date	2020-05-22	Test Completion Date	2020-06-05
Test Standard and Method	Technical Standard for Disinfection (2002 Ministry of Health P.R.China)- 2.1.5.4		
Item Tested	Validation of disinfection efficacy of sterilization and disinfection instruments		
Test Conclusion	<p>The test data of the sample(s) is attached to the page(s) of this report.</p> <p>Issue Date: 2020-06-16 (Official Seal)</p>		
Remarks	SL001,SL002,SL003 and SL004 are different from the main test sample, except for the model, the sample name and function are the same. (provided by the applicant)		

Editor:

Qiunanchun

Verifier:

Sun Jiusi

Approver:

Xie Xiaobao



GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No.: [REDACTED]

1. Sample pretreatment: The glass was used as the carrier of the microbial sheet and placed in the center of sterilizer for 10minutes.

2. Test Result:

Test germ	Test group	Average number of positive controls (cfu/PCS)	Average number of testing groups (cfu/PCS)	Killing rate (%)	Sterilization logarithm
<i>Escherichia coli</i> 8099	1	2.6×10^6	10	99.99	5.42
	2	2.6×10^6	10	99.99	5.42
	3	2.5×10^6	20	99.99	5.10
<i>Staphylococcus aureus</i> ATCC 6538	1	1.1×10^6	10	99.99	5.04
	2	1.2×10^6	10	99.99	5.08
	3	1.2×10^6	10	99.99	5.08
<i>Candida albicans</i> ATCC 10231	1	1.5×10^6	10	99.99	5.17
	2	1.6×10^6	20	99.99	4.90
	3	1.4×10^6	10	99.99	5.15

Sample Picture



(Blank below)



Report №.: [REDACTED]

Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.

广微测
检测中心
章